

Pyramid Projection - Validation of a New Method of Skin Defect Measurement



RŽIŽKA, J., NOVÝ, P., VÁVRA, F., BOLEK, L., BENEŠ, J. Pyramid Projection - Validation of a New Method of Skin Defect Measurement. *Physiological Research*, 2007, ro. 56, ? Suppl. 1, s. 99-105. ISSN: 0862-8408

This paper presents a new method for the determination of the volume, surface area and depth of skin defects. The method is based on the description of a spatial defect using a pyramid. The projection of the pyramid is photographed using a digital camera and defect is mathematically reconstructed.